

Nanoscale STED-microscopy with fluorescent proteins [Abstract]

Katrin Willig, Stefan Hell

Angaben zur Veröffentlichung / Publication details:

Willig, Katrin, and Stefan Hell. 2007. "Nanoscale STED-microscopy with fluorescent proteins [Abstract]." In *Focus on Microscopy, April 10th-13th, 2007, Valencia, Spain*. Augsburg: Universität Augsburg.

Nutzungsbedingungen / Terms of use:

licgercopyright

Dieses Dokument wird unter folgenden Bedingungen zur Verfügung gestellt: / This document is made available under these conditions:

Deutsches Urheberrecht

Weitere Informationen finden Sie unter: / For more information see:

<https://www.uni-augsburg.de/de/organisation/bibliothek/publizieren-zitieren-archivieren/publiz/>



NANOSCALE STED-MICROSCOPY WITH FLUORESCENT PROTEINS

Katrin I. Willig, Stefan W. Hell
Department of NanoBiophotonics
Max-Planck-Institute for Biophysical Chemistry
Am Fassberg 11, 37077 Göttingen, Germany
kwillig@gwdg.de

KEY WORDS: Stimulated emission depletion, STED, fluorescent protein

Stimulated emission depletion (STED)[1] is a concept to overcome the diffraction resolution barrier fundamentally: The excitation beam is overlapped with a second, red-shifted beam featuring a zero in the center that is capable of quenching excited molecules and thus reducing the focal spot to sub-diffraction sizes.

We report subdiffraction resolution using stimulated emission depletion (STED) microscopy with fluorescent protein labeled samples[2]. The ~70 nm lateral resolution attained in this study is demonstrated with images of GFP-labeled viruses and the endoplasmic reticulum of a mammalian cell. Our results enable nanoscale biological microscopy with genetically encoded markers.

[1] S. W. Hell, and J. Wichmann, "Breaking the diffraction resolution limit by stimulated emission." *Opt. Lett.* , **19**, 780-782 (1994)

[2] K. I. Willig, R. R. Kellner, R. Medda, B. Hein, S. Jakobs, and S. W. Hell, "Nanoscale resolution in GFP-based microscopy" *Nature Methods* **3**, 721-723 (2006).