

## **Transtemporal investigation of brain parenchyma elasticity using 2-D shear wave elastography: validation appreciated! [Letter]**

**Michael Ertl, Nele Raasch, Gertrud Hammel, Katharina Harter, Christopher Lang**

### **Angaben zur Veröffentlichung / Publication details:**

Ertl, Michael, Nele Raasch, Gertrud Hammel, Katharina Harter, and Christopher Lang. 2019. "Transtemporal investigation of brain parenchyma elasticity using 2-D shear wave elastography: validation appreciated! [Letter]." *Ultrasound in Medicine & Biology* 45 (5): 1344.  
<https://doi.org/10.1016/j.ultrasmedbio.2019.01.004>.

## TRANSTEMPORAL INVESTIGATION OF BRAIN PARENCHYMA ELASTICITY USING 2-D SHEAR WAVE ELASTOGRAPHY: VALIDATION APPRECIATED!

On behalf of all authors we appreciate the letter of Tzschätzsch et al. (2019) regarding potential discrepancies in magnetic resonance and sonography values. Indeed, we omitted a description of the method of elastography implemented with the Philipps iU22 system (Ertl et al. 2018). Based on a literature research, our system uses the shear modulus ( $G$ ) approach (Nguyen et al. 2014). Therefore, the discussion of discrepancies between the reported technique and magnetic resonance elastography methods also using the shear modulus would be obsolete. Nevertheless, this conversion would be a very good explanation for the discrepancies between the values generated in our population and those in the study by Chauvet et al. (2016).

In summary, our main aim was to investigate the feasibility of the ultrasound method and the data acquisition in normal brain tissue. As our data suggest, the values in normal brain tissue seem to be reliable, with relatively little variation between the probands. From a sonographic point of view, we therefore think that these data can serve as a basis for further studies of pathologic conditions. The thickness of the skull is obviously an issue, but in the presence of a sufficient bone window, we were able to generate consistent values.

Of course, validation of our data by other groups would be highly appreciated.

MICHAEL ERTL\*  
NELE RAASCH\*  
GERTRUD HAMMEL†  
KATHARINA HARTER†  
CHRISTOPHER LANG\*

\* Clinic for Neurology and Neurophysiology, Klinikum Augsburg, Augsburg, Germany; and † Chair and Institute of Environmental Medicine, UNIKA-T, Technical University of Munich and Helmholtz Zentrum München, Augsburg, Germany

Address correspondence to: Michael Ertl, Clinic for Neurology and Neurophysiology, Klinikum Augsburg, Stenglinstrasse 2, 86156 Augsburg, Germany.

E-mails: [michlertl@web.de](mailto:michlertl@web.de), [Michael.Ertl@klinikum-augsburg.de](mailto:Michael.Ertl@klinikum-augsburg.de)

## REFERENCES

- Chauvet D, Imbault M, Capelle L, Demene C, Mossad M, Karachi C, Boch AL, Gennisson JL, Tanter M. In vivo measurement of brain tumor elasticity using intraoperative shear wave elastography. *Ultraschall Med* 2016;37:584–590.
- Ertl M, Raasch N, Hammel G, Harter K, Lang C. Transtemporal investigation of brain parenchyma elasticity using 2-D shear wave elastography: Definition of age-matched normal values. *Ultrasound Med Biol* 2018;44:78–84.
- Nguyen MM, Zhou S, Robert JL, Shamdasani V, Xie H. Development of oil-in-gelatin phantoms for viscoelasticity measurement in ultrasound shear wave elastography. *Ultrasound Med Biol* 14;40:168–176.
- Tzschätzsch H, Kreft B, Braun J, Sack I. Transtemporal investigation of brain parenchyma elasticity using 2-D shear wave elastography: Trustworthy? *Ultrasound Med Biol* 2019;45 00–00.

<https://doi.org/10.1016/j.ultrasmedbio.2019.01.004>