

Monitoring cell productivity for the production of recombinant proteins by flow cytometry: an effective application using the cold capture assay


Katharina V. Meyer, Ina G. Siller, Jana Schellenberg, Alina Gonzalez Salcedo, Dörte Solle, Jens Matuszczyk, Thomas Scheper, Janina Bahnemann

Angaben zur Veröffentlichung / Publication details:

Meyer, Katharina V., Ina G. Siller, Jana Schellenberg, Alina Gonzalez Salcedo, Dörte Solle, Jens Matuszczyk, Thomas Scheper, and Janina Bahnemann. 2021. "Monitoring cell productivity for the production of recombinant proteins by flow cytometry: an effective application using the cold capture assay." *Engineering in Life Sciences* 21 (5): 288–93. <https://doi.org/10.1002/elsc.202000049>.

SHORT COMMUNICATION

Monitoring cell productivity for the production of recombinant proteins by flow cytometry: An effective application using the cold capture assay

Katharina V. Meyer¹  | Ina G. Siller¹ | Jana Schellenberg¹ |
Alina Gonzalez Salcedo¹ | Dörte Solle¹ | Jens Matuszczyk² | Thomas Scheper¹ |
Janina Bahnemann¹

¹ Institute of Technical Chemistry, Leibniz University Hannover, Hannover, Germany

² Sartorius Stedim Biotech GmbH, Göttingen, Germany

Correspondence

Dr. Janina Bahnemann, Institute of Technical Chemistry, Leibniz University Hannover, Callinstraße 5, 30167 Hannover, Germany.

Email: jbahnemann@iftc.uni-hannover.de

Funding information

Deutsche Forschungsgemeinschaft, Grant/Award Number: 346772917; Open Access fund of Leibniz Universität Hannover

Abstract

Due to the increasing economic and social relevance of biotherapeutics, their production processes are continually being reconsidered and reoptimized in an effort to secure higher product concentrations and qualities. Monitoring the productivity of cultured cells is therefore a critically important part of the cultivation process. Traditionally, this is achieved by determining the overall product titer by high performance liquid chromatography (HPLC), and then calculating the specific cell productivity based on this titer and an associated viable cell density. Unfortunately, this process is typically time-consuming and laborious. In this study, the productivity of Chinese Hamster Ovary (CHO) cells expressing a monoclonal antibody was analyzed over the course of the cultivation process. In addition to calculating the specific cell productivity based on the traditional product titer determined by HPLC analysis, culture productivity of single cells was also analyzed via flow cytometry using a cold capture assay. The cold capture assay is a cell surface labelling technique described by Brezinsky et al., which allows for the visualization of a product on the surface of the producing cell. The cell productivity results obtained via HPLC and the results of cold capture assay remained in great accordance over the whole cultivation process. Accordingly, our study demonstrates that the cold capture assay offers an interesting, comparatively time-effective, and potentially cheaper alternative for monitoring the productivity of a cell culture.

Abbreviations: a. u., arbitrary units; APC, allophycocyanin; CHO, Chinese Hamster Ovary; CVCT, cumulative viable cell time; ELISA, enzyme-linked immunosorbent assay; HPLC, high performance liquid chromatography; mAb, monoclonal antibody; MFI, mean fluorescence intensity; PBS, phosphate buffered saline; qP, specific cell productivity; SF, shake flask; VCD, viable cell density

This is an open access article under the terms of the [Creative Commons Attribution](https://creativecommons.org/licenses/by/4.0/) License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

© 2021 The Authors. *Engineering in Life Sciences* published by Wiley-VCH GmbH

KEYWORDS

CHO cells, mammalian cell culture, monoclonal antibody production, secretion assay, specific cell productivity

Biotherapeutics derived from mammalian cell cultures are constantly gaining importance within the pharmaceutical industry [2,3]. The increased demand for these products is continually driving a push for ever more efficient production processes [4–6]. Product yield of recombinant protein is essentially determined by two factors: The count of viable cells and the specific productivity of the cells during the cultivation process [7]. As a result, carefully and accurately monitoring these two parameters is of importance in developing a high-yielding process. The most common method for achieving this is to determine product titer via either a high performance liquid chromatography (HPLC) analysis [7–9] or an enzyme-linked immunosorbent assay (ELISA) [2,10–12]. However, both of these methods are regrettably time-consuming and laborious. For example, ELISA protocols involve several incubation steps and the addition of different reagents, which have to be handled manually [2,12]. HPLC systems require high maintenance and the measurements themselves are rather lengthy [13].

In this study, the productivity of Chinese Hamster Ovary (CHO) cells expressing a monoclonal antibody (mAb) was analyzed over the course of the cultivation process. In addition to calculating the specific cell productivity based on the product titer determined by HPLC analysis, the productivity was also analyzed by flow cytometry using the cold capture assay. A schematic overview of the analysis is shown in Figure 1. The cold capture assay was first described by Brezinsky et al. [1] and constitutes a cell surface labelling technique using fluorescently labelled antibodies that bind to the secreted target proteins. This method is based on the fact that proteins for secretion are transported from the Golgi apparatus to the cell membrane via vesicles that join the plasma membrane and so release their content into the extracellular environment [1,4]. Therefore proteins secreted by a cell can be visualized on its surface [1,4]. This method has previously been used to enrich high producing cells by fluorescent activated cell sorting [1,2,14,15], but to the best of the authors' knowledge, this study represents the first time that it has been reported as a method of successfully monitoring the productivity of cultured cells over the time course of the cultivation process.

A Cellca CHO DG44 cell line (Sartorius, DE) expressing a human IgG1 mAb was used for the experiments. The cultivation process started by conducting a seed culture. In brief, the cryoconserved cells were thawed and passaged

five times, every 3–4 days. For all passages single-use 0.5 L Erlenmeyer Shake Flasks (Corning, USA) with 150 mL pre-warmed cell line associated stock culture medium (Sartorius, DE) were used. The cells were incubated (Heracell™ 240, Thermo Fisher Scientific, USA) at a temperature of 36.8°C, 85% humidity, a pCO₂ of 7.5% and a shaking rate of 120 rpm with an orbital diameter of 19 mm. The resulting main culture was then inoculated with 0.3×10^6 cells·mL⁻¹ in three single-use 0.125 L Erlenmeyer Shake Flasks (Corning, USA), with 25 mL pre-warmed cell line associated basal medium for production (Sartorius, DE). The main culture was run as a fed-batch process, starting with a batch-phase of 72 h, for 12 cultivation days in total. Starting from day 5 of this process, a glucose solution (Sigma Aldrich, USA) was supplied in addition to the cell line associated daily feed (Sartorius, DE) once the




FIGURE 1 Flow chart of the performed experiments. The productivity of the cultured cells was analyzed based on VCD and product titer as well as by flow cytometry

PRACTICAL APPLICATION

Monitoring the productivity of cultured cells is an important tool for cultivation process optimization. This is frequently done by determining the product titer using HPLC analysis which can be a time-consuming and laborious process. The cold capture assay offers an interesting alternative, through which researchers can potentially monitor cell culture productivity using flow cytometry. This method offers segregated data about all populations instead of integral data such as the overall volumetric titer concentrations. Based on this high producer clones can be separated or the process conditions can be adjusted to a high production of all cells resulting in high titer concentrations.

glucose level dropped below $5 \text{ g}\cdot\text{L}^{-1}$. The main culture was kept at the same culture conditions described above for the seed culture, and all used feeds and media were chemically defined.

Throughout this experiments, a sample was taken daily from each shake flask (SF 1-3). Viable cell density (VCD) and viability of the culture were analyzed using a trypan blue assay-based cell counter (Cedex HiRes, Roche, CHE). The titer of the mAb in the cell culture supernatant was determined by a HPLC (Chromaster VWR International GmbH, USA) method with a size exclusion column (Yarra™ 3 μm SEC 3000 OOH-4513-KO 300 x 7.8 mm, Phenomenex Inc., USA) controlled by the HPLC-Software Open Lab Control Panel (Agilent Technologies, USA). The method was 20 min long with a flow rate of $1 \text{ mL}\cdot\text{min}^{-1}$, an oven temperature of 25°C and a sample injection volume of $5 \mu\text{L}$. The product peak was measured using a diode array detector (280 nm), and the product concentration was determined using a standard curve prepared with known product concentrations. For the HPLC analysis a buffer (pH 6.6) containing 100 mL of 1 M sodium sulfate (Carl Roth GmbH & Co. KG, DE), 50 mL 1 M sodium dihydrogen phosphate (Sigma-Aldrich, Merck KGaA, DE), 50 mL 1 M disodium hydrogen phosphate (Sigma-Aldrich, Merck KGaA, DE) and 800 mL ddH₂O (Arium® Sartorius, DE) was used.

Figure 2 shows the VCD, viability and mAb concentration of the main culture over the course of the cultivation process. The parameters demonstrated similar development across all three shake flasks. In all three flasks, cell viability remained above 90% up until and including day 10, and thereafter decreased to approximately 81% on the last cultivation day. As expected, the maximal VCD was

reached on day 8 (SF1: $12.44 \times 10^6 \text{ cells}\cdot\text{mL}^{-1}$, SF2: $14.99 \times 10^6 \text{ cells}\cdot\text{mL}^{-1}$, SF3: $14.75 \times 10^6 \text{ cells}\cdot\text{mL}^{-1}$). Until day 5, only a slight increase in product titer could be observed, while thereafter the titer showed a larger increase. The highest titer was reached on day 11 in shake flasks 1 and 3 (SF1: $2.1 \text{ g}\cdot\text{L}^{-1}$, SF3: $2.5 \text{ g}\cdot\text{L}^{-1}$), and on day 12 in shake flask 2 (SF2: $2.4 \text{ g}\cdot\text{L}^{-1}$). Shake flask 1 and Shake flask 3 also showed a decrease in product titer on day 12, at the end of the cultivation process. This decrease could have been caused by the decreasing viability at the end of the cultivation. However, it bears noting that the viability decreased similarly across all three shake flasks while the decrease of the mAb titer is only observed for two of the three shake flasks. VCD and product titer were used to calculate the specific productivity of the cells. Since cell viability and mAb titer decreased notably on the last day of cultivation, specific cell productivity was only calculated up until day 11. The specific cell productivity (qP) [$\text{pg}\cdot\text{cell}^{-1}\cdot\text{day}^{-1}$] was calculated employing the following Equation 1 as described in Edros, McDonnell and Al-Rubea [10]. Here, mAb represents the antibody concentration at a particular time while the cumulative viable cell time (CVCT) is calculated employing Equation 2, where x_n is the VCD [$10^6 \text{ cells}\cdot\text{mL}^{-1}$] at a particular time t_n [day] and x_{n+1} the VCD [$10^6 \text{ cells}\cdot\text{mL}^{-1}$] after the elapsed time t_{n+1} [day].

$$qP = \frac{mAb_{n+1} - mAb_n}{CVCT} \quad (1)$$

$$CVCT = \frac{x_{n+1} + x_n}{2} \times (t_{n+1} - t_n) \quad (2)$$

The calculated specific cell productivity of the main culture is shown in Figure 3 in blue. Its time course was similar for all three shake flasks. A maximum of the specific cell productivity was also observed across all three shake flasks on day 6 of the cultivation (SF1: $49.71 \text{ pg}\cdot\text{cell}^{-1}\cdot\text{day}^{-1}$, SF2: $43.33 \text{ pg}\cdot\text{cell}^{-1}\cdot\text{day}^{-1}$, SF3: $46.42 \text{ pg}\cdot\text{cell}^{-1}\cdot\text{day}^{-1}$). This maximum is in accordance with the already mentioned great increase of the product titer from day 5 to day 6. The other process related maximum of the specific cell productivity towards the end of the cultivation process (day 10/11) could also be observed.

In addition to calculating the specific cell productivity based on the product titer determined by HPLC analysis and VCD, the productivity of the main culture was also analyzed by flow cytometry using the cold capture assay. In this study a staining method based on the information provided by Brezinsky et al. [1] and Pichler et al. [4] regarding the cold capture assay was used and the manufacturer specifications of the utilized anti-human-IgG1-allophycocyanin (APC) antibody (Miltenyi Biotec, DE) were considered. Briefly, the cell suspension was




FIGURE 2 VCD, viability and mAb concentration during the cultivation in all shake flasks (SF1-3)




FIGURE 3 Specific cell productivity of the main culture calculated according to Equation 1 (blue) and percentage of cells in the cold capture assay (bright red) for all three shake flasks (SF1-3). As a scale of variation for the cold capture assay, the mean fluorescence intensity (MFI) in arbitrary units (a. u.) of the negative control is shown as well (dark red)

centrifuged at room temperature for 5 min at $300 \times g$ (Centrifuge 5702, Eppendorf, DE) and a pellet of up to 10^6 cells was then resuspended in $98 \mu\text{L}$ 4°C cold phosphate buffered saline (PBS). Subsequently, $2 \mu\text{L}$ of the staining antibody were added and the suspension was then mixed by careful pipetting and kept in the dark at 4°C for 15 min. Afterwards, these cells were washed using 1 mL of 4°C cold PBS and centrifuged at 4°C for 5 min at $300 \times g$ (Heraeus Megafuge 8R Thermo Fisher Scientific, USA). The supernatant was thereafter discarded, and the cell pellet was resuspended in $500 \mu\text{L}$ 4°C cold PBS and kept on ice. A BD Accuri™ C6 (Becton Dickinson, USA) was used to conduct the flow cytometry analysis. For all samples, at least 10,000

events were measured and data analysis was run using BD Accuri™ C6 software (Becton Dickinson, USA), R [16] and TinnR [17]. Unstained Cellca CHO DG44 cells were used as a negative control. In further experiments stained non-producing Cellca CHO DG44 cells should be included as an additional control. Intact cells were gated using the forward scatter-area vs. side scatter-area signal and a doublet discrimination was conducted. The corresponding red fluorescence signal of the APC linked to the anti-human-IgG1 antibody has an excitation maximum of 650 nm and an emission maximum of 660 nm [18] and was captured using appropriate filter settings. Finally, the percentage of events showing a higher red fluorescent signal in the side

scatter-area vs. APC-area plot than the negative control was analyzed.

The results of the cold capture assay are shown in Figure 3 in bright red. The time course of the signal obtained in the cold capture assay was similar for all three shake flasks. A maximum could be observed on day 6 of the cultivation (SF1: 39.04%, SF2: 35.48%, SF3: 41.04%), the same day that the calculated specific cell productivity reached its apex. Similar to the specific cell productivity, the signal obtained in the cold capture assay also showed an increase towards the end of the cultivation. Further experiments with a larger sample size could be considered for a statistical analysis of the correlation between the specific cell productivity and the results of the cold capture assay. Overall, the specific cell productivity determined via HPLC and the results of the cold capture assay remained in great accordance over the whole cultivation process. When comparing the two methods, however, it should be noted that the cell productivity determined by HPLC analysis is averaged over the entire cell population (integral data), whereas in the cold capture assay using flow cytometry the cells are analyzed individually (segregated data). Thus, the latter method offers a great potential for process optimization, since by observing individual cells, high producers can be identified and selected. It should also be considered that the calculation of the specific productivity via HPLC analysis is dependent on the product titer as well as the VCD. As a result, this method requires the input of data obtained from two instruments, while the cold capture assay results are based on data from only one instrument. Furthermore, the specific cell productivity calculated with the help of HPLC results is influenced by values previously obtained, since analysis of the supernatant inherently reflects previous processes and already secreted proteins. In contrast, the signal obtained by cold capture assay shows the current state of the cells. This may prove beneficial for monitoring quick changes of productivity during a cultivation process. The results demonstrate that the use of flow cytometry to monitor cell productivity therefore delivers a comparatively fast and reliable analysis, and consequently allows for comparatively quicker adaptations of ongoing production processes. Especially in situations where the titer determination via HPLC is complicated for a specific product, the cold capture assay offers an easier to use and more time efficient method to secure an overview of the productivity of a conducted culture process. Moreover, with only limited additional effort, the analysis of other culture-relevant parameters is also possible using this method for example, observing of the apoptosis/necrosis status of a culture. In summary, the cold capture assay offers an excellent alternative to more traditional methods for monitoring productivity during a cultivation

process and shows great potential for further process optimization.

ACKNOWLEDGMENTS

The authors acknowledge the financial support of the German Research Foundation (DFG) via the Emmy Noether Programme (346772917). Furthermore the authors would like to thank the Open Access fund of Leibniz Universität Hannover for the funding of the publication of this article.

Open access funding enabled and organized by Projekt DEAL.


CONFLICT OF INTEREST

The authors have declared no conflict of interest.

DATA AVAILABILITY STATEMENT

The data that support the findings of this study are available from the corresponding author upon reasonable request.

ORCID

Katharina V. Meyer  <https://orcid.org/0000-0003-1198-6664>

REFERENCES

1. Brezinsky, S. C. G., Chiang, G. G., Szilvasi, A., Mohan, S., et al., A simple method for enriching populations of transfected CHO cells for cells of higher specific productivity. *J. Immunol. Methods* 2003, 277, 141–155.
2. Pichler, J., Galosy, S., Mott, J., and Borth, N., Selection of CHO host cell subclones with increased specific antibody production rates by repeated cycles of transient transfection and cell sorting. *Biotechnol. Bioeng.* 2011, 108, 386–394.
3. Kumar, N. and Borth, N., Flow-cytometry and cell sorting: an efficient approach to investigate productivity and cell physiology in mammalian cell factories. *Methods* 2012, 56, 366–374.
4. Pichler, J., Hesse, F., Wieser, M., Kunert, R., et al., A study on the temperature dependency and time course of the cold capture antibody secretion assay. *J. Biotechnol.* 2009, 141, 80–83.
5. Fischer, S., Handrick, R., and Otte, K., The art of CHO cell engineering: a comprehensive retrospect and future perspectives. *Biotechnol. Adv.* 2015, 33, 1878–1896.
6. Tharmalingam, T., Barkhordarian, H., Tejeda, N., Daris, K., et al., Characterization of phenotypic and genotypic diversity in subclones derived from a clonal cell line. *Biotechnol. Prog.* 2018, 34, 613–623.
7. Du, Z., Treiber, D., Mccarter, J. D., Fomina-Yadlin, D., et al., Use of a small molecule cell cycle inhibitor to control cell growth and improve specific productivity and product quality of recombinant proteins in CHO cell cultures. *Biotechnol. Bioeng.* 2015, 112, 141–155.
8. Meleady, P., Doolan, P., Henry, M., Barron, N., et al., Sustained productivity in recombinant Chinese Hamster Ovary (CHO) cell lines: proteome analysis of the molecular basis for a process-related phenotype. *BMC Biotechnol.* 2011, 11, 78.

9. Clarke, C., Doolan, P., Barron, N., Meleady, P., et al., Predicting cell-specific productivity from CHO gene expression. *J. Biotechnol.* 2011, 151, 159–165.
10. Edros, R. Z., McDonnell, S., and Al-Rubeai, M., Using molecular markers to characterize productivity in Chinese hamster ovary cell lines. *PLoS One* 2013, 8, e75935.
11. Moretti, P., Behr, L., Walter, J. G., Kasper, C., et al., Characterization and improvement of cell line performance via flow cytometry and cell sorting. *Eng. Life Sci.* 2010, 10, 130–138.
12. Sleiman, R. J., Gray, P. P., McCall, M. N., Codamo, J., et al., Accelerated cell line development using two-color fluorescence activated cell sorting to select highly expressing antibody-producing clones. *Biotechnol. Bioeng.* 2008, 99, 578–587.
13. Boysen, R. I. and Hearn, M. T., HPLC of peptides and proteins: preparation and system set-up. *Curr. Protoc. Mol. Biol.* 2001, 54, 10.12.1–10.12.9.
14. Okumura, T., Masuda, K., Watanabe, K., Miyadai, K., et al., Efficient enrichment of high-producing recombinant Chinese hamster ovary cells for monoclonal antibody by flow cytometry. *J. Biosci. Bioeng.* 2015, 120, 340–346.
15. Gallagher, C. and Kelly, P. S., Selection of high-producing clones using FACS for CHO cell line development. in: Meleady, P. (Ed.). *Heterologous Protein Production in CHO Cells. Methods in Molecular Biology*. Vol. 1603. Humana Press, New York 2017, pp. 143–152.
16. R Core Team, R: A Language and Environment for Statistical Computing. R Foundation for Statistical Computing, Vienna, Austria 2020.
17. Faria, J. C., Resources of Tinn-R GUI/Editor for R Environment. UESC, Ilheus, Bahia, Brasil 2012.
18. Becton Dickinson and Company, *Absorption and Emission Spectra*. 2019. URL: <https://www.bdbiosciences.com/en-us/applications/research-applications/multicolor-flow-cytometry/product-selection-tools/spectrum-guide-page> (visited on 11/04/2019).

How to cite this article: Meyer KV, Siller IG, Schellenberg J, et al. Monitoring cell productivity for the production of recombinant proteins by flow cytometry: An effective application using the cold capture assay. *Eng Life Sci.* 2021;21:288–293. <https://doi.org/10.1002/elsc.202000049>